

STARTUP CLEANING PROCEDURE

Are you starting up?

(First user of the day on Weekends, Holidays and outside weekly working hours)

Turn on the Cytometer and the Computer
Login on Windows and Accuri C6 Plus software
Follow the **startup cleaning procedure**:



diH₂O
RUN + FAST
wait 15 min

You are ready to run your samples!

Thank you!